

Memo

date: January 30, 2014

to: Thomas Roser

from: Yousef Makdisi (on behalf of the Experiment Safety Review Committee)

*subject: Recommendation for approval to introduce flammable gas into the STAR
Detector for RUN 14.*

The STAR detector was subjected to an annual ESRC review on January 16, 2014. The PASS flammable gas detectors and IR ventilation fans have been certified by an email from Reich on January 15, 2014. The Blue Sheet items were completed and signed off on January 30, 2014. What remains is for CAS watch to carry out a handshake with the experiment on January 30 2014 prior to starting the flammable gas flow. The Walkthrough was carried out on January 29, 2014 and no action items relevant to this were found. The crane is locked out and a shift schedule is in place.

With this, approval is requested to introduce flammable gases into the STAR detectors.

All relevant documentation is in the C-AD Experiment Safety Review Committee files.

cc:

W. Christie
P. Ingrassia
Y. Makdisi
Xu Nu
A. Pendzick
P. Pile
B. Mueller

File

Approved: 

Date: 1/30/14