

## Memo

*date: January 12, 2016*


*to: Thomas Roser*

*from:  Yousef Makdisi (On behalf of the ESRC)*

*subject: Recommendation for approval to introduce flammable gas into the STAR  
Detector.*

The STAR detector was submitted to an Annual Safety Review on December 21, 2015. The Walkthrough was carried out on January 11, 2016. The PASS flammable gas detectors and IR ventilation fans have been certified and the Blue sheeting was completed and signed off. The crane is locked. A shift crew is in place.

With this, approval is requested to introduce flammable gas into the STAR detector systems.

Approved: 

Date: 1/12/16

**cc:**

B. Mueller  
W. Christie  
R. Pak  
C. Folz  
P. Pile  
~~X~~Xu

File