


BLIP Operations – 186 days of operation

- **Safety** – None.
- **Production**
 - Two processing went well - Arronax and BNL
 - Product shipped on the 20 June.
- **Maintenance** – None
- **Research**
 - Thorium target was scheduled to start on Fri 20 June.
 - **Failed helium test on 19 June** (late in the day) prevent irradiation on Friday.
 - **Alternative target holder** with reduced target material was prepared and safety review conducted at 2 pm on Friday.
 - Delayed irradiation for Monday and commence at 4.30 pm.

Current delivered to BLIP 18 Dec – 24 June



Total Integrated Current at BLIP vs Days of Operation

